

# BlueDot

## ANA<sup>10</sup> IgG

Order Code: ANA10D-24

### 1. INTENDED USE

BlueDot ANA<sup>10</sup> IgG is an immunodot kit intended for the detection, in human sera only, of IgG autoantibodies against the antigens Sm, Sm/RNP, SSA/Ro 60kD, SSB, Jo-1, Scl-70, PM-Scl 100, Ku, CENP-A/B and PCNA.

This kit is intended to confirm results of anti-nuclear patterns obtained by immunofluorescence, the screening and reference method in autoimmunity; the kit is intended as an aid in the diagnosis of several autoimmune diseases (for more details, see 11.5 *Autoantibodies diagnostic values*).

The test is intended for confirmation of IFA positive population and IFA negative with strong suspicion of autoimmune disorder.

This kit is strictly reserved for professional use in clinical analysis laboratories. Prior training is strongly recommended (please contact your distributor).

It can only be used manually on a platform shaker or in an open automated immunodot processing system, programmed according to the pipetting scheme described in point 9.2.

Detection of the different IgG autoantibodies can be either qualitative (see point 10.1) or semi-quantitative (see point 10.2).

### 2. PRINCIPLE OF THE TEST

This kit and all its components are intended for exclusive manual use.

The test is based on the principle of an Enzyme Immunoassay. The strips are composed of a membrane fixed onto a plastic support. In the assay procedure, the strips are incubated with diluted patient serum. If present in the sample, the patient's autoantibodies bind to the specific antigen on the membrane. Unbound or excess antibodies are removed by washing in the next step. Then, human anti-IgG immunoglobulins conjugated with alkaline phosphatase are incubated with the strips and bind to the antigen-antibody complexes on the membrane surface. After a second washing step to remove excess conjugate, the chromogen/substrate solution is added, resulting in the appearance of an insoluble coloured product (purple) that precipitates at the site of the enzymatic reaction. The colour intensity is directly proportional to the quantity of antibodies present in the sample. The kit contains 24 single-use tests.

### 3. KIT CONTENTS

**Prior to any use of the kit, please check that all the items listed are present. Please also check if the characteristics of the product are corresponding to those described hereafter. If one of the items is missing or damaged, please do not use the kit and contact your distributor.**

#### 3.1 COMPONENTS

|                   |                         |                                                                                                                                                      |                                                                                                                                                                                                                                                                                                                                             |
|-------------------|-------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| TO BE<br>DILUTED: | (10 x) Wash<br>Solution | 1 x 40 ml - 10x concentrated (colourless)<br>Contains: H <sub>2</sub> O • TBS • NaCl • Tween • Preservatives                                         |  <br><br>RC — ○<br>Sm — ○<br>Sm/RNP — ○<br>SSA/Ro 60kD — ○<br>SSB — ○<br>Jo-1 — ○<br>Scl-70 — ○<br>PM-Scl 100 — ○<br>Ku — ○<br>CENP-A/B — ○<br>PCNA — ○<br>CO — ○ |
| READY TO<br>USE:  | Dot strips              | 24 units (each strip is for single use)<br>12 dots each:<br>1 negative control (CO)<br>10 antigens<br>1 positive control (RC)                        |                                                                                                                                                                                                                                                                                                                                             |
|                   | Sample Diluent          | 1 x 40 ml (yellow)<br>Contains: H <sub>2</sub> O • TBS • NaCl • Tween • BSA • Preservatives • Dye                                                    |                                                                                                                                                                                                                                                                                                                                             |
|                   | Conjugate               | 1 x 40 ml (red)<br>Contains: H <sub>2</sub> O • TBS • NaCl • KCl • MgCl <sub>2</sub> • AP-conjugated goat anti-human IgG • Preservatives • Dye       |                                                                                                                                                                                                                                                                                                                                             |
|                   | Substrate               | 1 x 40 ml (brown bottle, pale yellow solution)<br>Contains: H <sub>2</sub> O • Preservatives • MgCl <sub>2</sub> • TBS • NBT • BCIP • NBT Stabilizer |                                                                                                                                                                                                                                                                                                                                             |
|                   | Incubation trays        | 3 units<br>with 8 wells for incubation                                                                                                               |                                                                                                                                                                                                                                                                                                                                             |

Abbreviations in alphabetic order:

AP = Alkaline Phosphatase; BCIP = Bromo-Chloro-Indolyl-Phosphate; BSA = Bovine Serum Albumin; KCl = Potassium Chloride; MgCl<sub>2</sub> = Magnesium Chloride; NaCl = Sodium Chloride; NBT = NitroBlue Tetrazolium; TBS = Tris Buffer Saline

For more information on the composition and concentration of the active ingredients used, please refer to the MSDS available on request or on [www.d-tek.be](http://www.d-tek.be).

### Symbols used on kit labels

|                                                                                   |                                    |                                                                                   |                                      |
|-----------------------------------------------------------------------------------|------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------|
|  | Consult instructions for use       |  | CE Mark + Notified Body              |
|  | In Vitro Diagnostic Medical Device |  | For 24 uses                          |
|  | To be stored from 2-8°C            |  | Reference                            |
|  | Batch Number                       |  | To be protected from direct sunlight |
|  | Use-by date                        |  | Manufacturer                         |
|  | Cartridge                          |  | Caution                              |
|  | Strip                              |                                                                                   |                                      |

### 3.2 Antigens used

|             |                                                                                                                                                             |
|-------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Sm          | Core proteins of snRNP particles. Contains mainly D protein. E, F, G subunits are detectable. BB' proteins are not detectable (purified from bovine thymus) |
| Sm/RNP      | snRNP particles. Contains essentially 68kD, A, BB', C and D proteins. A significant amount of snRNA is detectable (purified from bovine thymus)             |
| SSA/Ro 60kD | Ro 60 kD protein (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                                          |
| SSB         | La 50 kD protein (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                                          |
| Jo-1        | Histidyl-tRNA synthetase (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                                  |
| Scl-70      | DNA topoisomerase I (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                                       |
| PM-Scl 100  | Polymyositis-Scleroderma antigen (100 kD subunit) (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                         |
| Ku          | Regulatory subunit of DNA-dependent protein kinase (70/80 kD heterodimer) (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                 |
| CENP-A/B    | Centromere Proteins A and B (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                               |
| PCNA        | Proliferating Cell Nuclear Antigen (recombinant, human, expressed in Baculovirus-infected Sf9 cells)                                                        |

### 3.3 Reactive ingredients

| Substance                                | Origin                                                          | Intend purpose in ANA Screening kits                        | Concentration in ANA Screening kits                      | Purity                                                                     |
|------------------------------------------|-----------------------------------------------------------------|-------------------------------------------------------------|----------------------------------------------------------|----------------------------------------------------------------------------|
| Goat anti-human IgG-alkaline phosphatase | Animal (Goat)                                                   | Secondary antibody (detection antibody) in conjugate buffer | < 0,1 µg/ml in conjugate buffer                          | Unknown. No antibody is detectable to non-immunoglobulin serum components. |
| Sm                                       | purified from bovine thymus                                     | Biomarker (antigen) coated on strips                        | 0,04 mg/ml<br>One Sm spot = 0,5µl on each strip          | >80%                                                                       |
| Sm/RNP                                   | purified from bovine thymus                                     | Biomarker (antigen) coated on strips                        | 0,067 mg/ml<br>One Sm/RNP spot = 0,5µl on each strip     | >80%                                                                       |
| SSA/Ro 60kD                              | recombinant, human, expressed in Baculovirus-infected Sf9 cells | Biomarker (antigen) coated on strips                        | 0,01 mg/ml<br>One SSA/Ro 60kD spot = 0,5µl on each strip | >80%                                                                       |
| SSB                                      | recombinant, human, expressed in Baculovirus-infected Sf9 cells | Biomarker (antigen) coated on strips                        | 0,01 mg/ml<br>One SSB spot = 0,5µl on each strip         | >80%                                                                       |
| Jo-1                                     | recombinant, human, expressed in Baculovirus-infected Sf9 cells | Biomarker (antigen) coated on strips                        | 0,02 mg/ml<br>One Jo-1 spot = 0,5µl on each strip        | >80%                                                                       |
| Scl-70                                   | recombinant, human, expressed in Baculovirus-infected Sf9 cells | Biomarker (antigen) coated on strips                        | 0,02 mg/ml<br>One Scl-70 spot = 0,5µl on each strip      | >80%                                                                       |
| PM-Scl 100                               | recombinant, human, expressed in Baculovirus-infected Sf9 cells | Biomarker (antigen) coated on strips                        | 0,2 mg/ml<br>One PM-Scl 100 spot = 0,5µl on each strip   | >80%                                                                       |

|                                   |                                                                  |                                      |                                                                                     |          |
|-----------------------------------|------------------------------------------------------------------|--------------------------------------|-------------------------------------------------------------------------------------|----------|
| Ku                                | recombinant, human, expressed in Baculovirus- infected Sf9 cells | Biomarker (antigen) coated on strips | 0,025 mg/ml<br>One Ku spot = 0,5µl on each strip                                    | >80%     |
| CENP-A/B                          | recombinant, human, expressed in Baculovirus-infected Sf9 cells  | Biomarker (antigen) coated on strips | CEN-A: 0,033 mg/ml<br>CEN-B: 0,025 mg/ml<br>One CENP A/B spot = 0,5µl on each strip | >80%     |
| PCNA                              | recombinant, human, expressed in Baculovirus-infected Sf9 cells  | Biomarker (antigen) coated on strips | 0,2 mg/ml<br>One PCNA spot = 0,5µl on each strip                                    | >80%     |
| Protein L                         | Bacterial (from Peptostreptococcus magnus)                       | Reactive (positive) control          | 0,01 mg/ml<br>One RC spot = 0,5µl on each strip                                     | >95%     |
| Streptavidin-Alkaline Phosphatase | Bacterial (from Streptomyces avidinii)                           | Cut-off (negative) control           | < 0,1 µg/ml<br>One CO spot = 0,5µl on each strip                                    | Unknown. |
| NBT-BCIP                          | Synthetic (chemical substance)                                   | Substrate for alkaline phosphatase   | 0,2 mg/ml                                                                           | ≥ 98%    |

#### 4. MATERIAL REQUIRED BUT NOT PROVIDED

Platform shaker / Micropipettes / Timer / Graduated cylinder / Distilled or deionised water / Tweezers / Absorbent and/or filter paper.

#### 5. STORAGE

The reconstituted wash solution is stable for at least one month at 2-8°C. Reagents and strips can be stored at 2-8°C until the expiry date indicated on each vial or tube.

Place unused strips back into the provided tube, seal it and store at 2-8°C. Chromogen/Substrate (NBT/BCIP) shall be stored at 2-8°C.

When stored properly, all test kit components are stable until the indicated expiry date.

#### 6. SAFETY PRECAUTIONS

- All reagents are for in vitro diagnostic and professional use only. The test kit should be processed by trained technical staff only.
- The reagents in the kit are considered as not dangerous, as the concentrations of potentially dangerous chemicals are below the thresholds specified by European regulations:

| Name              | CAS       | EINECS | Concentration in strip | Classification according to Regulation EC 1272/2008<br>Significance<br>H Phrases |
|-------------------|-----------|--------|------------------------|----------------------------------------------------------------------------------|
| Cellulose Nitrate | 9004-70-0 | -      | < 5 %                  | Flam. Sol. 1<br>H228                                                             |

Annex VI to Regulation (EC) No 1272/2008: Index N°: 603-037-00-6; Commission Regulation (EU) 2015/830; 3.2.1

| Name | CAS        | EINECS | Concentration in mixture | Classification (in concentrated form) according to Regulation EC 1272/2008<br>Significance<br>H Phrases                                                                                                           |
|------|------------|--------|--------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| MIT: | 55965-84-9 | -      | < 0,0015 %               | Acute Tox. 2 H330<br>Acute Tox. 2 H310<br>Acute Tox. 3 H301<br>Skin Corr. 1 C H314; C ≥ 0,6%<br>Eye Dam. 1 H318; C ≥ 0,6%<br>Skin Sens. 1 A H317; C ≥ 0,0015%<br>A Aquatic Acute 1 H400<br>Aquatic Chronic 1 H410 |

Annex to Commission Regulation (EU) 2018/1480; Index Number: 613-167-00-5 ; Commission Regulation (EU) 2015/830; 3.2.1

| Name             | CAS        | EINECS    | Concentration in mixture | Classification (in concentrated form) according to Regulation EC 1272/2008<br>Significance<br>H Phrases                    |
|------------------|------------|-----------|--------------------------|----------------------------------------------------------------------------------------------------------------------------|
| NaN <sub>3</sub> | 26628-22-8 | 247-852-1 | < 0.1 %                  | Acute tox. 2<br>H300<br>Acute tox. 1<br>H310<br>STOT RE 2<br>H373<br>Aquatic acute 1<br>H400<br>Aquatic chronic, 1<br>H410 |

Annex VI to Regulation (EC) No 1272/2008: Index Number: 011-004-00-7; Commission Regulation (EU) 2015/830; 3.2.1

| Name | CAS      | EINECS    | Concentration in mixture | Classification (in concentrated form) according to Regulation EC 1272/2008<br>Significance<br>H Phrases |
|------|----------|-----------|--------------------------|---------------------------------------------------------------------------------------------------------|
| NBT  | 298-83-9 | 206-067-4 | < 0,01%                  | Acute tox. 4<br>H302                                                                                    |

Nevertheless, these chemicals are toxic in concentrated form. Therefore, contact with the skin, eyes or mucous membranes should be avoided by using suitable individual protection (gloves, laboratory coat, goggles). As with any chemical containing specific hazards, the product/components of the product should only be handled by qualified personnel and with the necessary precautions.

3. Patient samples should be handled as if they could transmit infectious diseases; they therefore require suitable protection (gloves, laboratory coat, goggles). In any case, GLP should be applied with all the general or individual safety rules in force.
4. Waste disposal: Patient samples, incubated test strips and used reagent vials should be handled as infectious waste. The boxes and other containers do not need to be collected separately, unless stated otherwise in official regulations.
5. The device is containing substances from animal, human, and bacterial origins (cfr 3.3) at very low concentration. All these substances were selected in order to not contain any microbial or transmissible agents and are non-toxic at the concentration used in the device. Nevertheless, laboratory good practice at user site (glasses, gloves) is necessary

## 7. RECOMMANDATIONS

1. D-tek and its authorized distributors cannot be held responsible for damages caused indirectly or due to: a change or modification in the indicated procedure, an improper use of the kit and / or the use of an incomplete or damaged kit. The use of this kit is reserved for qualified technical personnel only.
2. D-tek's responsibility is limited in all cases to the replacement of the kit.
3. In the event of a serious incident (injury, deterioration in health, or death) with this IVD device, please report it immediately to the manufacturer (see address below) and to the competent authority in your country.

## 8. SAMPLE COLLECTION, HANDLING AND STORAGE

Sera with particles should be centrifuged at low speed. Blood samples should be collected in dry tubes. Please avoid using a pool of different sera, as this can lead to inconsistent results (see point 10.4). After separation, the serum samples should be used immediately or aliquoted and stored at 2-8 ° C (for maximum 14 days) or frozen at -20°C (for longer storage periods, maximum 13 months). Repeated freezing/ thawing cycles of the samples must be at a maximum of 10 cycles.

## 9. ASSAY PROCEDURE

### **BASIC INFORMATION, HANDLING AND TIPS:**

The dots are precoloured blue on the strips, ensuring that all antigens have been dotted correctly onto the membrane. This blue coloration disappears during the first step of the incubation. During incubation with the wash solution, a faint pink background coloration appears on the membrane and disappears upon drying at the end of the procedure.

During the procedure, agitation of the incubation tray is necessary to ensure efficient circulation of fluids over the membrane. A Rocking platform is the shaker of choice. Be sure to adjust the movement of the shaker in such a way that no spilling of solutions or cross-contamination between the wells can occur.

After each filling of the wells with solution, agitate manually the incubation tray until the strips are completely immersed in order to remove air bubbles which may be trapped under the strip. Alternatively, floating strips may be forced into the solution by pushing down (with tweezers or pipette tip) on the upper part of the strip (plastic label zone).

**Avoid touching** the membrane zone of the strip with fingers, tweezers or pipette tips. Always use the plastic label zone for handling or manipulation. The whole procedure must be run **at room temperature (18-25°C)**.

### **Description of the CONTROLS:**

The **Positive Control or RC (Reaction Control)** consists of a protein (protein L) fixing all the immunoglobulins present in the test sample. If the test has been carried out correctly, this control will show a colouring at the end of the test (with an intensity depending on the effective concentration of immunoglobulins in the sample).

The absence of any colouring of this dot at the end of the test may indicate that the sample has not been pipetted on the strip (see 10.4 *Troubleshooting*).

The **Negative Control or CO (Cut-Off Control)** consists of a protein (streptavidin – alkaline phosphatase) reacting with the enzymatic substrate and with certain constituent elements of the tested sample. If the test has been carried out correctly, this control is coloured at the end of the test, with a signal depending on the kinetics of the substrate and the characteristics of the

sample. The intensity of this control serves as a threshold value for the final interpretation of the results (see 10 *INTERPRETATION OF RESULTS*).

### 9.1 Reagents preparation

1. Allow all components to equilibrate at room temperature (**18-25°C**) before use.
2. **Dilute** the concentrated **wash solution 10x** with **distilled water**.  
*Prepare 15 ml diluted wash solution per strip tested*  
*Example: 1,5 ml concentrated wash solution + 13,5 ml distilled water for one strip.*  
**Do not substitute reagents or mix strips with different batch numbers this may lead to variations in the results.**

### 9.2 Pipetting flow chart

1. **Place** one **strip** per patient into the wells, blue dots **facing up**.
2. Add **2 ml diluted wash solution** per well. **Incubate** (shake) **for 10 min**.  
*Upon correct incubation, the blue coloration of the dots completely disappears.*  
*If not prolong the procedure until the colour of the dots fades completely.*
3. **Discard** solution from the wells.  
*Remove liquid by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edge of the tray with absorbent paper.*
4. Add **1,5 ml sample diluent** per well.
5. Add **10 µl patient sample** per well. **Incubate** (shake) **for 30 min**.  
*Avoid touching the membrane with the pipette tip. Preferentially dispense the sample into the solution over the upper part of the strip (plastic label zone).*  
*Note: Steps 4 and 5 can be combined by pre-diluting the sample in a glass or plastic tube (1,5 ml sample diluent + 10 µl patient sample). Mix (Add to the well)*
6. **Discard** solution from the wells.  
*Remove liquid by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edge of the tray with absorbent paper.*
7. **Wash 3 x 3 minutes** with **1,5 ml diluted wash solution** per well (shake).  
*Following each wash step remove liquid from the wells by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edges of the tray with absorbent paper*
8. Add **1,5 ml Conjugate** per well. **Incubate** (shake) **for 30 min**.
9. **Discard** solution from the wells.  
*Remove liquid by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edge of the tray with absorbent paper*
10. **Wash 3 x 3 min.** with **1,5 ml diluted wash solution** (shake)  
*Following each wash step remove liquid from the wells by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edges of the tray with absorbent paper.*
11. Add **1,5 ml Substrate** per well. **Incubate** (shake) **for 10 min**.
12. **Discard** solution from the wells.  
*Remove liquid by slowly inverting the plate. The strips will adhere to the bottom of the wells. Dry the edge of the tray with absorbent paper.*
13. **Wash 1 x 3 min.** with **1,5 ml diluted wash solution** per well to stop the reaction.
14. **Collect** the strips from the wells and allow them to dry for 30 minutes on absorbent paper. The interpretation must be done in the 24 hours following the test processing.

## 10. INTERPRETATION OF RESULTATS

A visual (qualitative) interpretation of the results of the kit is possible, however the use of the BlueScan scanner and the Dr Dot Software is generally recommended for more precision and a semi-quantitative interpretation.

**IMPORTANT NOTICE: The positivity of all parameters of this kit is NOT possible and in such a case the test is not valid. An additional test must be performed to establish the diagnosis!**

### 10.1. Qualitative Interpretation

1. Peel off the cover of the adhesive on the back side of each strip and attach strips dots face up onto the marked fields of the interpretation sheet provided with the kit. This will indicate the respective positions of the different controls and antigens on the membrane.
2. The first upper dot (**Positive Control Dot**) must be positive for all patients. Only a clearly coloured Positive Control Dot ensures your results are valid and operation was correct and/or kit components were not degraded. If the first upper dot is not coloured, the test has failed and cannot be interpreted further.
3. Compare the specific **antigen** dots to the **Negative Control Dot** (which always is the last bottom dot). The colour intensity of the antigen dots is directly proportional to the titer of the specific antibody in the patient sample.  
*The colour intensity of the Negative Control Dot may vary depending on the sample characteristics. If the sample is free of interfering substances the Negative Control Dot may be even close to uncoloured. In contrast, a highly coloured Negative Control Dot indicates a high rate of unspecific binding in the sample.*

#### **POSITIVE RESULT:**

A sample is positive for a specific antibody if the colour intensity of the corresponding Antigen dot is higher than the intensity of the Negative Control Dot.

#### **NEGATIVE RESULT:**

A sample is negative for a specific antibody if the colour intensity of corresponding Antigen dot is lower than or equal to the intensity of the Negative Control Dot.

NB: A weak coloration of an antigen dot, when close to the colour intensity of the Negative Control Dot may be difficult to differentiate by visual inspection only. In such cases, it is recommended to use Dr Dot Software and scanning system (see 10.2) and refer to the corresponding instructions for more accurate interpretation.

## 10.2 Results semi-quantification: use of Dr Dot Software and Scanning system (material needed: BlueDiver Clamp, empty stripholders)

The BlueScan scanner is a specially designed system for the reading of D-tek immunodot strips. It allows precise and easy insertion of test strips.

The Dr Dot Software allows a semi-quantification of results. Based on the image obtained, each result will be quantified in grayscale value and compared to the reference scale integrated in the BlueScan Cover.

These grayscale intensities will be transformed and displayed in arbitrary units (AU, from 0 to 100) based on the intensities of the controls (RC and CO, see point 9) present on the strip, according to the following conversion formula:

$$\text{Result of antigen X (AU)} = \frac{\text{Grayscale intensity of antigen X} - \text{Grayscale intensity of CO}}{\text{Grayscale intensity of RC} - \text{Grayscale intensity of CO}} * 100$$

1. Prepare a BlueDiver Clamp and load it with as many empty stripholders as there are strips to analyse. Carefully insert a strip into each stripholder, RC showing upwards.
2. Insert the clamp, the reactive side of the strips facing down, into the dedicated emplacement in the cover of the BlueScan scanner.
3. Start scanning the strips using the Dr Dot Software.
4. The software semi-quantifies the results, and the interpretation of the obtained values is as follows

| Dr Dot arbitrary unit (AU) | Interpretation |
|----------------------------|----------------|
| < 5                        | Negative       |
| 5 – 10                     | Equivocal (*)  |
| > 10                       | Positive       |

For detailed information about the BlueScan and Dr Dot Software please refer to the Manual of Use of your Dr Dot Software

## 10.3 Important recommendations for the interpretation of results

1. D-tek's kits constitute a diagnostic aid. In consequence, no diagnosis can be established solely on the basis of our kits. The results should always be interpreted by taking into account the clinical examination, the patient's history and the results obtained by other methods.  
No single technique can rule out the possibility of false positive or false negative results. With this in mind, an indirect immunofluorescence test should, as far as possible, be carried out prior to the use of an immunodot kit (immunofluorescence being recognized as a reference method in autoimmunity).
2. The intensity of a result is not necessarily related to the degree of intensity of the disease, but rather to the level of antibodies detected.
3. Low titers of autoantibodies may occur in healthy patients. For this reason, low positive results (close to the CO, between 5 and 10 Dr Dot AU), although valid, should be considered equivocal. In such cases, the retesting of the patient, preferably by using a new sample, is recommended. If the result remains equivocal on retesting, other diagnostic tests and/or clinical information should be used to help determine the autoimmune status of the patient.
4. For various reasons, and under certain conditions, the kit may show a defect in performance (see 10.4 Troubleshooting). In such cases, the results are not valid and cannot be interpreted. It is recommended to repeat the test. If the error persists, please contact your distributor.
5. The intensity of the results may decrease when the device is used at the end of its life. However, the performance of the kit is not affected (detection of positives and negatives) under normal conditions of use and storage.
6. Sequential sampling (at different dates) of an autoimmune patient can sometimes lead to different results from one sample to another. This difference can have several reasons: the patient's treatment, the evolution of the disease, or a seroconversion. In the specific case of seroconversion, the result can be positive for an autoantibody in an early sampling of the patient and become positive for another autoantibody in a later sampling of the same patient.

## 10.4 Troubleshooting

| Problem                                                          | Possible causes + Action                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Discrepancy of results as compared to a reference method         | <div>- Use</div> <ul style="list-style-type: none"> <li>- incorrect pipetting of serum</li> <li>- incorrect volume dispensed</li> <li>- Use of two different samples of the same patient (see point 10.3.6) or wrong sample handling/storage between tests</li> <li>- erroneous visual interpretation</li> <li>- erroneous Dr Dot reading</li> </ul> <div>→ repeat the test</div> <div>- Material</div> <ul style="list-style-type: none"> <li>- Interfering substance in the sample</li> <li>- Sample is a pool of different human sera</li> </ul> <div>→ repeat the test and confirm by other methods</div> <div>- Method</div> <ul style="list-style-type: none"> <li>- intrinsic performance of the kit (see 11.2 Analytical sensitivity and specificity)</li> <li>- expired kit</li> <li>- stability problem</li> </ul> <p><b>Please contact your distributor for any further technical support requests.</b></p> |
| Different results in the same batch or between several batches - | <div>- Use</div> <ul style="list-style-type: none"> <li>- incorrect pipetting of serum</li> <li>- incorrect volume dispensed</li> <li>- erroneous visual interpretation or</li> <li>- bad Dr Dot reading</li> </ul>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    |

|                                                         |                                                                                                                                                                                                                                                                                                                                                                                                       |
|---------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                                                         | <p>→ repeat the test</p> <p>- Method      - intrinsic performance of the kit (see 11.1 <i>Repeatability and Reproducibility</i>)</p>                                                                                                                                                                                                                                                                  |
| Contamination between neighbouring strips               | <p>- Use      - incorrect pipetting of serum<br/>→ repeat the test</p>                                                                                                                                                                                                                                                                                                                                |
| RC absent or weak                                       | <p>- Use      - Serum not pipetted at all<br/>→ repeat the test<br/>- Patient with immunoglobulin deficiency<br/>→ repeat the test to confirm patient status<br/>- Damaged reagents<br/>→ check the integrity of the reagents<br/>→ contact your supplier if you suspect a problem<br/>- Spot not on the strip<br/>→ count the number of dots on the strip; if not correct, contact your supplier</p> |
| CO absent                                               | <p>- damaged reagents<br/>→ check the integrity of the reagents, contact your distributor if you suspect a problem<br/>- Spot absent from the strip<br/>→ count the number of spots present on the strip, contact your distributor in case of incorrect number</p>                                                                                                                                    |
| Non-specific bindings / high background / high CO value | <p>Suspected presence of a contaminant or an interfering substance in the patient sample<br/>→ repeat the test and confirm through another method</p> <p><b>Please contact your distributor for any further technical support requests.</b></p>                                                                                                                                                       |
| Strips not correctly labelled                           | Manufacturing problem → please contact your distributor                                                                                                                                                                                                                                                                                                                                               |
| Kit content incorrect                                   | Manufacturing problem → please contact your distributor                                                                                                                                                                                                                                                                                                                                               |
| Positive results for all the biomarkers of the kit      | Problem with reagents → please contact your distributor                                                                                                                                                                                                                                                                                                                                               |

#### NOTE:

The major residual risks of the kit, as given in the risk analysis of the kit at the end of design (after mitigation), are the following:

- 1) Risk of false results based on a pipetting error (bad serum)
- 2) Risk of false results based on an interfering substance contained in the sample

### 11. PERFORMANCES

#### 11.1 Repeatability and Reproducibility

Reference samples were tested for each antibody in successive statistically representative series, both in the same test as in different tests and between different batches in order to calculate the intra-assay, inter-assay and inter-lot variations respectively. In all the cases, the variations in colour intensity were within the following expected limits:

CV ≤ 10% for intra-assay runs  
CV ≤ 15% for inter-assay runs  
CV ≤ 20% for inter-lot runs

#### 11.2 Analytical sensitivity

Measurement range (semi-quantified results): From 0 AU (negative) to 100 AU (high positive).

Limit of detection: the lowest measured value of the test is 5 AU (considered as equivocal following the interpretation algorithm, see point 10.2)

As not any international standard is available for the autoantibodies, trueness of measurement and linearity are not applicable on this product.

#### 11.3 Analytical specificity

1. The main known interfering substances were tested on each biomarker of the present kit. For each concentration of interfering substance tested, the difference between the result of the sample without the interfering substance and the result obtained in the presence of the interfering substance did not exceed 15%.

| Interfering substance | Maximum Concentration | Intermediate Concentration | Minimum Concentration | Difference <15% |
|-----------------------|-----------------------|----------------------------|-----------------------|-----------------|
| Bilirubin             | 100 mg/dL             | 50 mg/dL                   | 25 mg/dL              | Yes             |
| Haemoglobin           | 200 mg/dL             | 100 mg/dL                  | 50 mg/dL              | Yes             |
| Cholesterol           | 224.3 mg/dL           | 112 mg/dL                  | 56 mg/dL              | Yes             |
| Rheumatoid factor IgM | ~500IU/ml             | ~300IU/ml                  | ~100IU/ml             | Yes             |

2. Note: It is impossible to test all the possible interfering substances described in the literature. Other interferences, amongst others drug-induced interferences, are possible. The high analytical specificity of the test is guaranteed by the quality of the antigen used. This kit detects IgG antibodies against Sm, Sm/RNP, SSA/Ro 60kD, SSB, Jo-1, Scl-70, PM-Scl 100, Ku, CENP-A/B and PCNA. No cross reactions with other autoantibodies have been found.

### 11.4 Clinical sensitivity and specificity

Sensitivity and Specificity were calculated from combined results obtained on clinically defined positive and negative EQAS controls and from historical data (external clinical evaluation on clinically defined positive and negative patients). These characterized samples (confirmed positive or negative for specific antibodies by reference laboratories and/or methodologies) were assayed following the test instructions. A detailed clinical report is available upon request.

| <b>Sensitivity:</b><br>The percentage is established with the following calculation:<br>$\text{Sensitivity} = \frac{\text{True Positive Results}}{\text{True Positive Results} + \text{False Negative Results}}$ |                       |                        |                 | <b>Specificity:</b><br>The percentage is established with the following calculation:<br>$\text{Specificity} = \frac{\text{True Negative Results}}{\text{True Negative Results} + \text{False Positive Results}}$ |                       |                        |                 |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------|------------------------|-----------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------|------------------------|-----------------|
| Antigen                                                                                                                                                                                                          | True Positive Results | False Negative Results | Sensitivity (%) | Antigen                                                                                                                                                                                                          | True negative results | False positive results | Specificity (%) |
| Sm                                                                                                                                                                                                               | 23                    | 0                      | >99             | Sm                                                                                                                                                                                                               | 244                   | 0                      | >99             |
| Sm/RNP                                                                                                                                                                                                           | 38                    | 1                      | 97              | Sm/RNP                                                                                                                                                                                                           | 175                   | 3                      | 98              |
| SSA/Ro 60kD                                                                                                                                                                                                      | 86                    | 0                      | >99             | SSA/Ro 60kD                                                                                                                                                                                                      | 132                   | 0                      | >99             |
| SSB                                                                                                                                                                                                              | 44                    | 2                      | 96              | SSB                                                                                                                                                                                                              | 172                   | 0                      | >99             |
| Jo-1                                                                                                                                                                                                             | 57                    | 0                      | >99             | Jo-1                                                                                                                                                                                                             | 162                   | 0                      | >99             |
| Scl-70                                                                                                                                                                                                           | 13                    | 0                      | >99             | Scl-70                                                                                                                                                                                                           | 206                   | 0                      | >99             |
| PM-Scl 100                                                                                                                                                                                                       | 3                     | 0                      | >99             | PM-Scl 100                                                                                                                                                                                                       | 31                    | 0                      | >99             |
| Ku                                                                                                                                                                                                               | 22                    | 2                      | 92              | Ku                                                                                                                                                                                                               | 26                    | 0                      | >99             |
| CENP-A/B                                                                                                                                                                                                         | 14                    | 1                      | 93              | CENP-A/B                                                                                                                                                                                                         | 200                   | 4                      | 98              |
| PCNA                                                                                                                                                                                                             | 13                    | 0                      | >99             | PCNA                                                                                                                                                                                                             | 42                    | 0                      | >99             |

Note: Sensitivity and specificity values of 100 % are strictly related to sample cohorts used in clinical evaluations. In theory, a diagnostic kit shouldn't be 100% sensitive or specific (at least > 99%).

### 11.5 Autoantibodies diagnostic values

|                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
|------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Anti-Sm          | Diagnostic marker (ACR and SLICC criterion) for systemic lupus erythematosus (SLE)<br>Diagnostic specificity of 99% for systemic lupus erythematosus (SLE)<br>Diagnostic sensitivity of 5-40 % for systemic lupus erythematosus (SLE)                                                                                                                                                                                                                                                                                                                                |
| Anti-Sm/RNP      | Sm :<br>Diagnostic marker (ACR and SLICC criterion) for systemic lupus erythematosus (SLE)<br>Diagnostic specificity of 99% for systemic lupus erythematosus (SLE)<br>Diagnostic sensitivity of 5-40 % for systemic lupus erythematosus (SLE)<br>RNP 68kD/A/C :<br>Diagnostic criterion of Mixed connective tissue disease (MCTD). Highly specific and extremely sensitive (100%) in the absence of Sm and dsDNA antibodies.<br>Found in 13 à 32 % of patients with systemic lupus erythematosus (SLE),<br>Found in 10 % of patients with systemic sclerosis (SSc)   |
| Anti-SSA/Ro 60kD | Diagnostic marker and classification criterion for Sjögren's Syndrome (SjS).<br>By EIA:<br>Found in 96% of patients with primary SjS,<br>Found in 80% of patients with secondary SjS<br>Found in 25-60% of patients with systemic lupus erythematosus (SLE),<br>Found in 90-100 % of patients with subacute cutaneous lupus erythematosus (SCLE)<br>Found in 90% of patients with neonatal cutaneous lupus erythematosus (NLE)<br>Found more rarely (5-15%) in patients with rheumatoid arthritis (RA) and<br>Found in 9% of patients with systemic sclerosis (SSc). |
| Anti-SSB         | Diagnostic marker for Sjögren's Syndrome (SjS)<br>By EIA:<br>Found in 70% of patients with primary SjS,<br>Found in 50% of patients with secondary SjS<br>Found in 25% of patients with systemic lupus erythematosus (SLE),<br>Found in 80% of patients with subacute cutaneous lupus erythematosus (SCLE)<br>Found in 70% of patients with neonatal cutaneous lupus erythematosus (NLE)                                                                                                                                                                             |
| Anti-Jo-1        | Diagnostic marker for idiopathic (autoimmune) myositis.<br>Diagnostic specificity of 100%, diagnostic sensitivity of 24-30% for auto-immune idiopathic myositis.                                                                                                                                                                                                                                                                                                                                                                                                     |
| Anti-Scl-70      | Diagnostic marker for Systemic Sclerosis (SSc)<br>Diagnostic specificity of 99%, sensibility of 10 % for limited SSc and up to 65% for diffuse SSc.                                                                                                                                                                                                                                                                                                                                                                                                                  |
| Anti-PM-Scl 100  | Diagnostic marker for connective tissue diseases with myositis and symptoms of systemic sclerosis.<br>Diagnostic specificity of 50-70% for polymyositis/scleroderma overlap syndrome, of 20% for idiopathic myositis and of 10% for systemic Sclerosis (SSc).<br>Diagnostic sensitivity of 24-55% for polymyositis/scleroderma overlap syndrome, of 8-12% in patients with idiopathic myositis and of 1-16% for systemic Sclerosis (SSc).                                                                                                                            |
| Anti-Ku          | Found in 23% of patients with "primary" pulmonary hypertension<br>Found in 1.8 à 23% of patients with systemic lupus erythematosus (SLE).<br>Found in 1.2 à 14 % of patients with systemic sclerosis (SSc).<br>Found in 2 à 33% of patients with an overlap syndrome with myositis.                                                                                                                                                                                                                                                                                  |
| Anti-CENP-A/B    | Diagnostic marker for Systemic Sclerosis (SSc).<br>Sensitivity of 57-82 % for patients with CREST syndrome (or other limited cutaneous forms of SSc) and in 3-12% of patients with diffuse cutaneous forms of SSc.                                                                                                                                                                                                                                                                                                                                                   |

|           |                                                                                  |
|-----------|----------------------------------------------------------------------------------|
|           | Detectable in 10-30 % of patients with Primary biliary cirrhosis (PBC).          |
| Anti-PCNA | Highly specific for systemic lupus erythematosus (SLE), but rarely found (3-7%). |

#### Publication references:

- 1: Orme ME, Andalusia C, Sjölander S, Bossuyt X. A comparison of a fluorescence enzyme immunoassay versus indirect immunofluorescence for initial screening of connective tissue diseases: Systematic literature review and meta-analysis of diagnostic test accuracy studies. *Best Pract Res Clin Rheumatol*. 2018 Aug;32(4):521-534. doi: 10.1016/j.berh.2019.03.005. Epub 2019 Apr 15. PMID: 31174821.
- 2: Jeong S, Hwang H, Roh J, Shim JE, Kim J, Kim GT, Tag HS, Kim HS. Evaluation of an Automated Screening Assay, Compared to Indirect Immunofluorescence, an Extractable Nuclear Antigen Assay, and a Line Immunoassay in a Large Cohort of Asian Patients with Antinuclear Antibody-Associated Rheumatoid Diseases: A Multicenter Retrospective Study. *J Immunol Res*. 2018 May 2;2018:9094217. doi: 10.1155/2018/9094217. PMID: 29854849; PMCID: PMC5954951.
- 3: Shovman O, Gilburd B, Chayat C, Amital H, Langevitz P, Watad A, Guy A, Perez D, Azoulay D, Blank M, Segal Y, Bentow C, Mahler M, Shoenfeld Y. Prevalence of anti-DFS70 antibodies in patients with and without systemic autoimmune rheumatic diseases. *Clin Exp Rheumatol*. 2018 Jan-Feb;36(1):121-126. Epub 2017 Jul 27. PMID: 28770702.
- 4: Zheng B, Wang Z, Mora RA, Liu A, Li C, Liu D, Zhai F, Liu H, Gong H, Zhou J, Liu J, Chen L, Wu L, Yuan L, Ying L, Jie L, He M, Hao M, Xu P, Lu Q, Han S, Chen S, Chen S, Zhu S, Sun W, Guo X, Chen Y, Wang Y, Qu Y, Li Z, Niu Z, Han Z, Chan EKL. Anti-DFS70 Antibodies Among Patient and Healthy Population Cohorts in China: Results From a Multicenter Training Program Showing Spontaneous Abortion and Pediatric Systemic Autoimmune Rheumatic Diseases Are Common in Anti-DFS70 Positive Patients. *Front Immunol*. 2020 Oct 2;11:562138. doi: 10.3389/fimmu.2020.562138. PMID: 33133072; PMCID: PMC7566153.
- 5: Hayashi N, Uto K, Imanishi A, Sugiyama D, Morinobu A, Saegusa J. Prevalence of anti-dense fine speckled 70 antibodies in healthy individuals and patients with antinuclear antibody-associated autoimmune rheumatic diseases in Japan. *Medicine (Baltimore)*. 2021 Mar 5;100(9):e24556. doi: 10.1097/MD.00000000000024556. PMID: 33655922; PMCID: PMC7939200.
- 6: Aberle T, Bourn RL, Munroe ME, Chen H, Roberts VC, Guthridge JM, Bean K, Robertson JM, Sivits KL, Rasmussen A, Liles M, Merrill JT, Harley JB, Olsen NJ, Karp DR, James JA. Clinical and Serologic Features in Patients With Incomplete Lupus Classification Versus Systemic Lupus Erythematosus Patients and Controls. *Arthritis Care Res (Hoboken)*. 2017 Dec;69(12):1780-1788. doi: 10.1002/acr.23201. Epub 2017 Nov 14. PMID: 28118528; PMCID: PMC5524597.
- 7: Zian Z, Maamar M, Aouni ME, Barakat A, Naima Ghailani Nourouti, El Aouad R, Arji N, Bennani Mechita M. Immunological and Clinical Characteristics of Systemic Lupus Erythematosus: A Series from Morocco. *Biomed Res Int*. 2018 Sep 30;2018:3139404. doi: 10.1155/2018/3139404. PMID: 30363993; PMCID: PMC6186365.
- 8: Wei Q, Jiang Y, Xiao M, Zhang X, Qi J, Xie J, Wu J, Wu Z, Gu J. Comparison of chemiluminescence microparticle immunoassay, indirect immunofluorescence assay, linear immunoassay and multiple microbead immunoassay detecting autoantibodies in systemic lupus erythematosus. *Scand J Immunol*. 2020 Mar;91(3):e12849. doi: 10.1111/sji.12849. Epub 2020 Jan 3. PMID: 31899559.
- 9: Au EY, Ip WK, Lau CS, Chan YT. Evaluation of a multiplex flow immunoassay versus conventional assays in detecting autoantibodies in systemic lupus erythematosus. *Hong Kong Med J*. 2018 Jun;24(3):261-269. doi: 10.12809/hkmj177007. Epub 2018 May 25. PMID: 29807953.
- 10: Betteridge ZE, Woodhead F, Lu H, Shaddick G, Bunn CC, Denton CP, Abraham DJ, du Bois RM, Lewis M, Wells AU, McHugh NJ. Brief Report: Anti-Eukaryotic Initiation Factor 2B Autoantibodies Are Associated With Interstitial Lung Disease in Patients With Systemic Sclerosis. *Arthritis Rheumatol*. 2016 Nov;68(11):2778-2783. doi: 10.1002/art.39755. PMID: 27273608.
- 11: René Louis Humbel, Groupe d'étude de l'auto-immunité (GEAI), l'info n°7, Mise au point anticorps anti Mi-2, Anticorps anti-DFS70/LEDGF/P75, p3, p6 mai 2015
- 12: Karsten Conrad, Werner Schössler, Falk Hiepe, Marvin J. Fritzler, Book "Autoantibodies in systemic Autoimmune Diseases", Volume 2, third edition – 2015

## 12. TEST LIMITATIONS

1. The results obtained with this confirmatory test are dependent on the intrinsic performance of the kit and must be considered as an aid to the final diagnosis, taking into account the results obtained by reference technique and the clinical data of the patient.
2. In case of hyper-lipemic samples, it is recommended to centrifuge it before the pipetting of the 10µl of sample, which must be done into the supernatant.
3. The concentration of autoantibodies in a serum sample is not relative to the results provided by the device.
4. There is no link between the concentration of the different autoantibodies detected by the device, and the severity of its associated autoimmune diseases.



We Apply Science



IFU – Instructions for Use  
ANA10D-24/p. 10 of 12



We Apply Science



IFU – Instructions for Use  
ANA10D-24/p. 11 of 12



We Apply Science



IFU – Instructions for Use  
ANA10D-24/p. 12 of 12